

Protocol for Striated Muscle Tissue Homogenization in the Bullet Blender™

The protocol described in this document is for the use of the Bullet Blender™ for the homogenization of striated / skeletal muscle tissue (from a variety of animals). Note that the time and speed settings may differ due to the variation in consistency/texture of tissue from species to species. This protocol does not specify a particular buffer - you may choose which is most appropriate for your downstream application (nucleic acid isolation, protein extraction, etc.).

Materials Required: muscle tissue, Bullet Blender™, homogenization buffer, pipettor, microcentrifuge tubes, and [0.5mm zirconium oxide beads \(part number ZSB05\)](#) or [0.5mm stainless steel beads \(part number ZrOB05\)](#).

Instructions

1. Cut muscle tissue into appropriately sized pieces for analysis (50mg-300mg) and place into a microcentrifuge tube. Typical sample size: 100mg.
2. **OPTIONAL:** Wash tissue 3x with ~1mL PBS. Aspirate. **NOTE:** This step removes external contaminants (blood, etc.).
3. Add zirconium silicate beads (0.5mm) **OR** zirconium oxide beads (0.5mm) to the tube. Use a mass of beads equal to 3x your mass of tissue. As a rough approximation, you may use half a volume of beads for each volume of tissue.
4. Add 2 volumes of buffer for every mass of tissue (for example: with 100mg tissue, add 0.2ml of buffer).
5. Close the microcentrifuge tubes.
6. Place tubes into the Bullet Blender™.
7. Set controls for **SPEED 9** and **TIME 3** minutes. Press **Start**.
8. After the run, remove tubes from the instrument.
9. Visually inspect samples. If homogenization is unsatisfactory, run for another two minutes at the **SPEED 10**.
10. Proceed with your downstream application.

SAFETY NOTE!!!

When using a centrifuge to separate your homogenate from the debris and beads, make sure your tubes are balanced.



Scientific Instrument Services, Inc.™

1027 Old York Rd. Ringoes, NJ 08551-1039

Phone: (908)788-5550

www.sisweb.com

Fax: (908) 806-6631