

# Electron Multiplier Cross Reference

A55

Man.	Mass Spec Model	Man. Part No.	Style	Galileo/Burle	Burle***	DeTech	ETP K & M**
Thermo* Elemental VG Elemental	Plasma Quad PQ1, PQ2+, PQ3 Standard			4870V	30140	401A	
Thermo Onix VG Gas Analysis	RGA (analog)					381	
Ulvac	Analog					378	
	Pulse Counting			4831G	30129	395	
UTI (MKS)	Analog - 100C	100C	Unmounted	4717	30074	101	
	Analog (mounted) - 100 C			4776	30102	333	
	Pulse Counting RGA - 100 C		Mounted	4876	30144	411	
	Discrete Dynode (Cube) - 100 C					415174-G1	
Vacuum Tech.	Aerovac RGA			4720	30075	392	
Varian	Auger Analyzer			4720	30075	312	
	Saturn I & II		Unmounted	4755	32782	102N-D	
	Saturn I, I, III		Mounted	4765G	30088	350	
	1200 Triple Quad			5904 Magnum	33328	2313	
	Saturn (pre 2000)		Mounted	4765G	30088		14138
	Saturn 2000, 2100, 2200			4765GM	32754	382	14147
	Saturn 4000 GC/MS			4765	32782	382	
	Ultramass						14566
	500					382	
Vestec	201 (Off Axis)			4782	30105	342	
	201 (On Axis) + ion mode			4783	33871	352	
V & F Gas Analysis	CIMS 500			4823G	30121	470	
VG Quad.	Quad Mass Spec (analog)	102-400-880		4774	30099	301	
	Quad Mass Spec (Pulse Counting)					401	
	QX200, SQ300, SQ400			4771	30095	304	
	SX200, SX300, SQ300	MGR4073B		4774	30099	301	
	ZABIF, 7070, ZAB			4773G	30098	R322	14130
	VG Smart IQ, VGQ			4769	30090		
VG Microtech	Clam2 Hemispherical Analyzer		Single Channel Detection	7010	30150M	226-10-C-LWL	
	Clam 2 Hemispherical Analyzer		Triple Channel Detection	4800B1 4X9 Cone	30106	224-4X19-C-LWL	
	Clam IV Hemispherical analyzer		Nine Channel Detection			244	
VG Scientific	ESCA Lab (single), ESCA Lab Mark II (Triple)			7010	30150M	226-10-C-LWL	
	ESCA Lab Mark II (five)			7013C/B1	30155	235-C-LWL	
	ESCA Single Probe (five), Micro Lab (five)					244	
VG Elemental	Genesis			4870V	30140		14568H
	PQ-3			4870V	30140		14214
	PQ-Excel			4870V	30140		
VG	X919VL			7010	30150M	226-10-C-LWL	
	X919BL			7010M	30151		
	B310VL			7012	33872	226-15-C-LWL	
	X812VL			7013C/B1	30155	235-C-LWL	
	X913VL			4800 C B1 4X19	30106	224-4X19-C-LWL	
Viking Instr.	Viking 573					2300	
VSW	Mass Analyst Quad Type			4771	30095	304	
Waters	Integrity MS					1368	

When "Thermo" is indicated, it is intended to imply "Thermo Fisher Scientific Corp." \*\*When "K & M" is indicated, it is intended to imply "ITT Power Solutions"

\*\*\*When "Burle" is indicated, it is intended to imply "Photonis USA, Inc."