

## A8 Other Mass Spectrometer Source Cleaning

### Perkin Elmer Inner Source Cleaning



#### Parts Cleaned

Entire source block, lenses, insulators, filament and repeller assembly.

#### Parts Repaired


Filament

#### Not included in standard source cleaning price

Replacement of missing screws, filaments and cracked, broken or missing insulators.

Part No.	Description	Price
REP-350	Source Cleaning of Perkin Elmer Inner Source	

### Perkin Elmer Outer Source Cleaning



#### Parts Cleaned

Entire source block, lenses, insulators.

#### Parts Replaced


All 4 ceramic rods are replaced with new ceramic rods

#### Not included in standard source cleaning price

Replacement of missing screws, copper spring contacts, other metal parts and cracked missing or broken insulators.

Part No.	Description	Price
REP-349C	Source Cleaning of Perkin Elmer Outer Source (includes replacement of Ceramics)	

### Nuclide 12-90-HT Source Cleaning



Part No.	Description	Price
REP-77	Source Cleaning of Nuclide 12-90-HT	

### MKS PVD Ion Source



Part No.	Description	Price
REP-75	Source Cleaning of MKS PVD Ion Source	

### Balzers QMA-400 Source Cleaning



Part No.	Description	Price
REP-75	Source Cleaning of Balzers QMA-400	

### Additional Mass Spectrometer Sources Repaired

In addition to the sources listed in the next few pages, SIS™ can clean and repair sources from any other mass spectrometers. Call SIS with your requirements. We will clean and repair your mass spectrometer source.